

**Certificate of Analysis  
GM501 SpermStore**

Batchnumber	SPSTORE-1824
Production date	18 December 2018
Expiry date	30 June 2020
Total volume (bottles)	880
Batchnumber HSA	B3AKC00053

	<b>Results</b>	<b>Specifications</b>
pH	7,46	7,20-7,60
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Endotoxin	< 0,05	< 0,25 EU/mL
Sperm survival Test 4 hours exposure	99	≥ 80 %

*Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.*

Release date 07 January 2019

Released by Dr. Julia Heinzmann

Signature

*p.p. Julia Heinzmann*