

**Certificate of Analysis
GM501 Bromelain in Dulbecco's PBS**

Batchnumber	BROM-2401
Production date	19 June 2024
Expiry date	31 March 2025
Total volume (ml)	2700

	Results	Specifications
pH	7,28	7,20-7,50
Osmolality	283	270-290 mOsm/kg
Endotoxin level	< 0,05	< 0,50 EU/mL
MEA (% blastocysts after 24 hours exposure)	100	≥ 80 %
MEA (% blastocysts after 96 hours exposure)	90,48	≥ 80 %
Sterility test (performed according to the current Ph. Eur. 2.6.1 /USP <71>)	No growth	No growth
HSSA (% motility compared with control after 4h exposure to test medium)	93	≥ 80 %

Release date	05 July 2024
Released by	Nicole Geiger
Signature	

pp. N. Geiger

REV 08_00 / 2023-10-30