

Certificate of Analysis GM501 GentleVit Thaw

Batchnumber	GT-2218
Production date	20 September 2022 / 22 September 2022
Expiry date	30 September 2023
Batchnumber HSA	B03G002753
Total volume (Kits)	200

	Results	Specifications
GM501 GentleVit Thaw 1		
pH	7,33	7,20-7,40
Sterility test	PASSED	PASSED
Endotoxin	< 0,1	< 0,25 EU/mL
GM501 GentleVit Thaw 2		
pH	7,31	7,20-7,40
Sterility test	PASSED	PASSED
Endotoxin	< 0,1	< 0,25 EU/mL
GM501 GentleVit Thaw 3		
pH	7,3	7,20-7,40
Sterility test	PASSED	PASSED
Endotoxin	< 0,1	< 0,25 EU/mL
Osmolality	840	805-850 mOsm/kg
GM501 GentleVit Thaw 4		
pH	7,35	7,20-7,40
Sterility test	PASSED	PASSED
Endotoxin	0,103	< 0,25 EU/mL
Osmolality	560	535-565 mOsm/kg
GM501 GentleVit Thaw 5		
pH	7,35	7,20-7,40
Sterility test	PASSED	PASSED
Endotoxin	< 0,1	< 0,25 EU/mL
Osmolality	427	405-435 mOsm/kg
GM501 GentleVit Thaw 6		
pH	7,32	7,20-7,40
Sterility test	PASSED	PASSED
Endotoxin	0,113	< 0,25 EU/mL
Osmolality	283	270-290 mOsm/kg
Mouse Embryo Assay (1-cell)	86	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	17. Oktober 2022
Released by	Vanessa Bierlich
Signature	