

Certificate of Analysis GM501 GentleVit Thaw

Batchnumber	GT-2224 + GT-2224A
Production date	27 December 2022
Expiry date	31 December 2023
Batchnumber HSA	B03G002753
Total volume (Kits)	161

	Results	Specifications
GM501 GentleVit Thaw 1		
pH	7,33	7,20-7,40
Sterility test	PASSED	PASSED
Endotoxin	< 0,1	< 0,25 EU/mL
GM501 GentleVit Thaw 2		
pH	7,29	7,20-7,40
Sterility test	PASSED	PASSED
Endotoxin	< 0,1	< 0,25 EU/mL
GM501 GentleVit Thaw 3		
pH	7,25	7,20-7,40
Sterility test	PASSED	PASSED
Endotoxin	< 0,1	< 0,25 EU/mL
Osmolality	841	805-850 mOsm/kg
GM501 GentleVit Thaw 4		
pH	7,31	7,20-7,40
Sterility test	PASSED	PASSED
Endotoxin	< 0,2	< 0,25 EU/mL
Osmolality	554	535-565 mOsm/kg
GM501 GentleVit Thaw 5		
pH	7,3	7,20-7,40
Sterility test	PASSED	PASSED
Endotoxin	< 0,1	< 0,25 EU/mL
Osmolality	421	405-435 mOsm/kg
GM501 GentleVit Thaw 6		
pH	7,32	7,20-7,40
Sterility test	PASSED	PASSED
Endotoxin	< 0,1	< 0,25 EU/mL
Osmolality	285	270-290 mOsm/kg
Mouse Embryo Assay (1-cell)	95	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

Release date	23. Februar 2023
Released by	Nicole Geiger
Signature	