

Certificate of Analysis GM501 GentleVit Thaw

Batchnumber	GT-2402
Production date	22 January 2024
Expiry date	31 January 2025
Total batch size	200
Batchnumber HSA	B03H025323


Parameter	Results	Specifications
GM501 GentleVit Thaw 1		
pH	7,33	7,20-7,40
Sterility test ¹	No growth	No growth
Endotoxin	< 0,2	< 0,25 EU/mL
Concentration HSA	16,4	13,9-20,9 g/l
GM501 GentleVit Thaw 2		
pH	7,3	7,20-7,40
Sterility test ¹	No growth	No growth
Endotoxin	< 0,2	< 0,25 EU/mL
Concentration HSA	17,2	14,4-21,6 g/l
GM501 GentleVit Thaw 3		
pH	7,34	7,20-7,40
Sterility test ¹	No growth	No growth
Endotoxin	< 0,2	< 0,25 EU/mL
Osmolality	833	805-850 mOsm/kg
Concentration HSA	17,4	14,9-22,3 g/l
GM501 GentleVit Thaw 4		
pH	7,34	7,20-7,40
Sterility test ¹	No growth	No growth
Endotoxin	< 0,23	< 0,25 EU/mL
Osmolality	553	535-565 mOsm/kg
Concentration HSA	19,1	15,4-23,1 g/l
GM501 GentleVit Thaw 5		
pH	7,35	7,20-7,40
Sterility test ¹	No growth	No growth
Endotoxin	< 0,22	< 0,25 EU/mL
Osmolality	417	405-435 mOsm/kg
Concentration HSA	19,6	15,7-23,6 g/l
GM501 GentleVit Thaw 6		
pH	7,37	7,20-7,40
Sterility test ¹	No growth	No growth
Endotoxin	< 0,2	< 0,25 EU/mL
Osmolality	283	270-290 mOsm/kg
Concentration HSA	19,9	16,0-24,0 g/l
Mouse Embryo Assay (1-cell)	95,24	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

¹Sterility test (according to the current PH. Eur. 2.6.1 / USP <71>)

Release date
Released by
Signature

8. Februar 2024
Nicole Geiger



REV 02_00 / 2023-09-15