

**Certificate of Analysis
GM501 GentleVit Thaw**

Batchnumber	GT-2408
Production date	02 April 2024
Expiry date	30 April 2025
Total batch size	200
Batchnumber HSA	B03H025323

Parameter	Results	Specifications
GM501 GentleVit Thaw 1		
pH	7,24	7,20-7,40
Sterility test ¹	NO GROWTH	No growth
Endotoxin	< 0,1	< 0,25 EU/mL
Concentration HSA	16,7	13,9-20,9 g/l
GM501 GentleVit Thaw 2		
pH	7,21	7,20-7,40
Sterility test ¹	NO GROWTH	No growth
Endotoxin	< 0,1	< 0,25 EU/mL
Concentration HSA	17,0	14,4-21,6 g/l
GM501 GentleVit Thaw 3		
pH	7,23	7,20-7,40
Sterility test ¹	NO GROWTH	No growth
Endotoxin	< 0,1	< 0,25 EU/mL
Osmolality	833	805-850 mOsm/kg
Concentration HSA	17,7	14,9-22,3 g/l
GM501 GentleVit Thaw 4		
pH	7,23	7,20-7,40
Sterility test ¹	NO GROWTH	No growth
Endotoxin	< 0,1	< 0,25 EU/mL
Osmolality	555	535-565 mOsm/kg
Concentration HSA	18,9	15,4-23,1 g/l
GM501 GentleVit Thaw 5		
pH	7,25	7,20-7,40
Sterility test ¹	NO GROWTH	No growth
Endotoxin	< 0,1	< 0,25 EU/mL
Osmolality	419	405-435 mOsm/kg
Concentration HSA	19,7	15,7-23,6 g/l
GM501 GentleVit Thaw 6		
pH	7,27	7,20-7,40
Sterility test ¹	NO GROWTH	No growth
Endotoxin	< 0,1	< 0,25 EU/mL
Osmolality	281	270-290 mOsm/kg
Concentration HSA	20,3	16,0-24,0 g/l
Mouse Embryo Assay (1-cell)	95,24	≥ 80 %

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.

¹Sterility test (according to the current PH. Eur. 2.6.1 / USP <71>)


Release date

24. April 2024

Released by

Nicole Geiger

Signature



REV 02_00 / 2023-09-15