

**Certificate of Analysis
Denu-Tips**

Batchnumber 44180203
 Expiry date 30 April 2021

Article No.	No. of Tips (vials)
GY-135/20	130
GY-145/20	90
GY-155/20	140
GY-170/20	120
GY-200/20	140

	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay <small>(1-cell, blastocysts after 96 hours)</small>	90,48	≥ 80 %

Release date 20 July 2018

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann