

**Certificate of Analysis
Denu-Tips**

Batchnumber 44180403
 Expiry date 31 October 2021

Article No.	No. of Tips (vials)
GY-135/xx	110
GY-140/xx	30
GY-145/xx	30
GY-150/xx	100
GY-170/xx	120
GY-175/xx	40
GY-200/xx	100
GY-275/xx	120
GY-300/xx	40

	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay <small>(1-cell, blastocysts after 96 hours)</small>	95,24	≥ 80 %

Release date 15 January 2019

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann