

**Certificate of Analysis
Denu-Tips**

Batchnumber 44190103
 Expiry date 31 January 2022

Article No.	No. of Tips (vials)
GY-140/20	200
GY-150/20	40
GY-135/5	20
GY-140/5	60
GY-145/5	20
GY-155/5	100
GY-165/5	60
GY-170/5	20
GY-275/5	20

	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay <small>(1-cell, blastocysts after 96 hours)</small>	95,24	≥ 80 %

Release date 12 February 2019
 Released by Dr. Julia Heinzmann
 Signature

p.p. Julia Heinzmann