

**Certificate of Analysis  
Denu-Tips**

Batchnumber 44190203  
Expiry date 30 April 2022

**Article No.**

- GY-135/xx
- GY-140/xx
- GY-145/xx
- GY-150/xx
- GY-165/xx
- GY-170/xx
- GY-175/xx
- GY-200/xx
- GY-275/xx
- GY-300/xx
- GY-550/xx

	<b>Results</b>	<b>Specifications</b>
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	95,24	≥ 80 %

Release date 30 April 2019  
Released by Dr. Julia Heinzmann

Signature

*p.p. Julia Heinzmann*