

**Certificate of Analysis  
Denu-Tips**

Batchnumber

44190203

Expiry date

30 April 2022

**Article No.**

GY-130/xx

GY-135/xx

GY-140/xx

GY-145/xx

GY-150/xx

GY-165/xx

GY-170/xx

GY-175/xx

GY-200/xx

GY-275/xx

GY-300/xx

GY-550/xx

**Results**

**Specifications**

Packaging

PASSED

PASSED

Sterilization

PASSED

PASSED

Mouse Embryo Assay

95,24

≥ 80 %

(1-cell, blastocysts after 96 hours)

Release date

17 June 2019

Released by

Dr. Julia Heinzmann

Signature

*p.p. Julia Heinzmann*