

**Certificate of Analysis  
Denu-Tips**

Batchnumber

44190403

Expiry date

31 October 2022

**Article No.**

GY-130/xx

GY-135/xx

GY-140/xx

GY-145/xx

GY-150/xx

GY-165/xx

GY-170/xx

GY-175/xx

GY-200/xx

GY-275/xx

GY-300/xx

GY-550/xx

**Result**

**Specification**

Mouse Embryo Assay  
(1-cell, blastocysts after 96 hours)

100

≥ 80 %

Release date

28 October 2019

Released by

Dr. Julia Heinzmann

Signature

*p.p. Julia Heinzmann*