

**Certificate of Analysis
Denu-Tips**

Batchnumber

44190403/A

Expiry date

31 October 2022

Article No.

GY-130/xx

GY-135/xx

GY-140/xx

GY-145/xx

GY-150/xx

GY-165/xx

GY-170/xx

GY-175/xx

GY-200/xx

GY-275/xx

GY-300/xx

GY-550/xx

Result

Specification

Mouse Embryo Assay
(1-cell, blastocysts after 96 hours)

90,48

≥ 80 %

Release date

22 November 2019

Released by

Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann