

**Certificate of Analysis
Denu-Tips**

Batchnumber 44200103/A

Expiry date 31 January 2023

Article No.

- GY-130/xx
- GY-135/xx
- GY-140/xx
- GY-145/xx
- GY-150/xx
- GY-155/xx
- GY-165/xx
- GY-170/xx
- GY-175/xx
- GY-200/xx
- GY-275/xx
- GY-290/xx
- GY-300/xx
- GY-550/xx

Result

Specification

<p>Mouse Embryo Assay (1-cell, blastocysts after 96 hours)</p>	<p>85,71</p>	<p>≥ 80 %</p>
---	--------------	---------------

Release date 14 February 2020

Released by Signature Dr. Julia Heinzmann

p.p. Julia Heinzmann

REV 05_00 / 2020-01-27