

Certificate of Analysis
Denu-Tips

Batchnumber	44200303/A
Production date	06 July 2020
Expiry date	31 July 2023

Article No.

GY-125/xx
 GY-130/xx
 GY-135/xx
 GY-140/xx
 GY-145/xx
 GY-150/xx
 GY-155/xx
 GY-165/xx
 GY-170/xx
 GY-175/xx
 GY-200/xx
 GY-275/xx
 GY-290/xx
 GY-300/xx
 GY-550/xx

	Result	Specification
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	90,48	≥ 80 %

Release date	04 August 2020
Released by Signature	Dr. Julia Heinzmann

p.p. Julia Heinzmann

REV 06_00 / 2020-07-01