

Certificate of Analysis
Denu-Tips

Batchnumber 44210403
 Production date 13 October 2021
 Expiry date 30 October 2024

Article No.

- GY-125/xx
- GY-130/xx
- GY-135/xx
- GY-140/xx
- GY-145/xx
- GY-150/xx
- GY-155/xx
- GY-165/xx
- GY-170/xx
- GY-175/xx
- GY-200/xx
- GY-275/xx
- GY-290/xx
- GY-300/xx
- GY-550/xx

	Result	Specification
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	95,24	≥ 80 %

Release date 19 October 2021

Released by Signature Dr. Julia Heinzmann

p.p. Julia Heinzmann

REV 06_00 / 2020-07-01