

**Certificate of Analysis
Denu-Tips**

Batchnumber	44220403
Production date	10 October 2022
Expiry date	31 October 2025

Article No.

GY-125/xx
 GY-130/xx
 GY-135/xx
 GY-140/xx
 GY-145/xx
 GY-150/xx
 GY-155/xx
 GY-165/xx
 GY-170/xx
 GY-175/xx
 GY-200/xx
 GY-275/xx
 GY-290/xx
 GY-300/xx
 GY-550/xx

	Result	Specification
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date	04 November 2022
Released by	Nicole Geiger
Signature	

pp. N. Geiger

REV 07_00 / 2022-09-06