

Certificate of Analysis
Denu-Tips

Batchnumber 442303/35/03
 Production date 28 August 2023
 Expiry date 31 August 2026

Article No.

GY-275/20

Parameter	Result	Specification
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	90	≥ 80 %

Release date 24 October 2023
 Released by Nicole Geiger
 Signature



REV 08_00 / 2023-09-15