

**Certificate of Analysis**  
**Denu-Tips**


Batchnumber 442401/08/01  
 Production date 20 February 2024  
 Expiry date 31 January 2027

**Article No.**

GY-200/20

<b>Parameter</b>	<b>Result</b>	<b>Specification</b>
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	95	≥ 80 %

Release date 25 April 2024  
 Released by Nicole Geiger  
 Signature



REV 08\_00 / 2023-09-15