

Certificate of Analysis
Denu-Tips


Batchnumber 442401/08/03
 Production date 20 February 2024
 Expiry date 31 January 2027

Article No.

GY-150/20

Parameter	Result	Specification
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date 25 April 2024
 Released by Nicole Geiger
 Signature



REV 08_00 / 2023-09-15