

**Certificate of Analysis
Denu-Tips**

Batchnumber 442403/33/02
 Production date 15 August 2024
 Expiry date 31 July 2027

Article No.

GY-150/20

Parameter	Result	Specification
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	95	≥ 80 %

Release date 14 October 2024
 Released by Nicole Geiger
 Signature



REV 08_00 / 2023-09-15