

**Certificate of Analysis  
Denu-Tips**

Batchnumber 442404/40/01  
 Production date 30 September 2024  
 Expiry date 30 September 2027

**Article No.**

GY-135/20

<b>Parameter</b>	<b>Result</b>	<b>Specification</b>
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date 03 December 2024  
 Released by Nicole Geiger  
 Signature

pp. N. Geiger

REV 08\_00 / 2023-09-15