

**Certificate of Analysis
Denu-Tips**


Batchnumber 442404/41/02
 Production date 10 October 2024
 Expiry date 30 September 2027

Article No.

GY-130/20

Parameter	Result	Specification
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	95	≥ 80 %

Release date 03 December 2024
 Released by Signature Nicole Geiger



REV 08_00 / 2023-09-15