

Certificate of Analysis
Denu-Tips

Batchnumber 442502/16/03
Production date 15 April 2025
Expiry date 31 March 2028

Article No.

GY-170/20

Parameter	Result	Specification
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	95	≥ 80 %

Release date 10 July 2025
Released by Nicole Geiger
Signature

pp. P. Geiger

REV 08_00 / 2023-09-15