

Certificate of Analysis
Denu-Tips

Batchnumber	442502/20/05
Production date	15 May 2025
Expiry date	30 April 2028

Article No.

GY-275/20

Parameter	Result	Specification
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date	09 July 2025
Released by	Nicole Geiger
Signature	

pp. N. Geiger

REV 08_00 / 2023-09-15