

Certificate of Analysis
Denu-Tips

Batchnumber 442503/31/02
Production date 01 August 2025
Expiry date 31 July 2028

Article No.

GY-135/20

Parameter	Result	Specification
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date 09 September 2025
Released by Nicole Geiger
Signature

pp. N. Geiger

REV 08_00 / 2023-09-15