

**Certificate of Analysis
Denu-Tips**

Batchnumber 442601/04/01
 Production date 19 January 2026
 Expiry date 31 December 2028

Article No.

GY-275/20

| Parameter | Result | Specification |
|--|---------------|----------------------|
| Packaging | PASSED | PASSED |
| Sterilization | PASSED | PASSED |
| Mouse Embryo Assay (1-cell, blastocysts after 96 hours) | 100 | ≥ 80 % |

Release date 25 February 2026
 Released by Dr. René Richarz
 Signature



REV 08_00 / 2023-09-15