

Certificate of Analysis
Denu-Tips

Batchnumber 442602/19/01
Production date 04 May 2026
Expiry date 30 April 2029

Article No.

GY-300/20

Parameter	Result	Specification
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date 08 July 2026
Released by Nicole Geiger
Signature

pp. N. Geiger

REV 08_00 / 2023-09-15