

Certificate of Analysis
Denu-Tips

Batchnumber	442602/22/02
Production date	25 May 2026
Expiry date	30 April 2029

Article No.

GY-275/20

Parameter	Result	Specification
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date	11 August 2026
Released by	Dr. René Richarz
Signature	

REV 08_00 / 2023-09-15