

Certificate of Analysis
Denu-Tips

Batchnumber 442602/22/03
Production date 25 May 2026
Expiry date 30 April 2029

Article No.

GY-150/20

Parameter	Result	Specification
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date 11 August 2026
Released by Dr. René Richarz
Signature



REV 08_00 / 2023-09-15