

**Certificate of Analysis
GM501 Collagenase**

Batchnumber	COLL-1905
Production date	09 October 2019
Expiry date	31 July 2020
Total volume (mL)	1130

	Results	Specifications
pH	7,46	7,10-7,50
Osmolality	280	270-310 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Sperm survival Test 4 hours exposure	95	≥ 80 %

Release date	24 October 2019
Released by	Dr. Julia Heinzmann
Signature	

p.p. Julia Heinzmann