

**Certificate of Analysis  
GM501 Collagenase**

Batchnumber	COLL-2001
Production date	13 January 2020
Expiry date	31 October 2020
Total volume (mL)	909

	<b>Results</b>	<b>Specifications</b>
pH	7,4	7,20-7,50
Osmolality	282	270-310 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Sperm survival Test 4 hours exposure	96	≥ 80 %

Release date	28 January 2020
Released by	Dr. Julia Heinzmann
Signature	

*p.p. Julia Heinzmann*