

Certificate of Analysis GM501 Collagenase

Batchnumber	COLL-2004
Production date	24 July 2020
Expiry date	30 April 2021
Total volume (mL)	1000

	Results	Specifications
pH	7,3	7,20-7,50
Osmolality	279	270-310 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Sperm survival Test 4 hours exposure	96	≥ 80 %

Release date	10 August 2020
Released by	Dr. Julia Heinzmann
Signature	

p.p. Julia Heinzmann