

**Certificate of Analysis
GM501 Collagenase**

Batchnumber	COLL-2104
Production date	24 March 2021
Expiry date	31 December 2021
Total volume (mL)	840

	Results	Specifications
pH	7,36	7,20-7,70
Osmolality	279	270-310 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Sperm survival Test 4 hours exposure	98	≥ 80 %

Release date	08 April 2021
Released by	Dr. Julia Heinzmann
Signature	

p.p. Julia Heinzmann

REV 09_00 /2021-02-08