

**Certificate of Analysis
GM501 Collagenase**

Batchnumber	COLL-2106
Production date	02 July 2021
Expiry date	30 April 2022
Total volume (mL)	840

	Results	Specifications
pH	7,36	7,20-7,70
Osmolality	278	270-310 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Sperm survival Test 4 hours exposure	89	≥ 80 %

Release date	19 July 2021
Released by	Vanessa Bierlich
Signature	

P.P. 

REV 09_00 /2021-02-08