

**Certificate of Analysis
GM501 Collagenase**

Batchnumber	COLL-2107
Production date	18 August 2021
Expiry date	31 May 2022
Total volume (mL)	1000

	Results	Specifications
pH	7,34	7,20-7,70
Osmolality	278	270-310 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Sperm survival Test 4 hours exposure	92	≥ 80 %

Release date	02 September 2021
Released by	Vanessa Bierlich
Signature	

P.P. 

REV 09_00 /2021-02-08