

**Certificate of Analysis  
GM501 Collagenase**

Batchnumber	COLL-2109
Production date	03 December 2021
Expiry date	30 September 2022
Total volume (mL)	840

	<b>Results</b>	<b>Specifications</b>
pH	7,35	7,20-7,70
Osmolality	282	270-310 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Sperm survival Test 4 hours exposure	96	≥ 80 %

Release date	21 December 2021
Released by	Dr. Julia Heinzmann
Signature	

*p.p. Julia Heinzmann*

REV 09\_00 /2021-02-08