

**Certificate of Analysis
GM501 Collagenase**

Batchnumber	COLL-2202
Production date	21 January 2022
Expiry date	31 October 2022
Total volume (mL)	1330

	Results	Specifications
pH	7,38	7,20-7,70
Osmolality	280	270-310 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Sperm survival Test 4 hours exposure	94	≥ 80 %

Release date	07 February 2022
Released by	Dr. Julia Heinzmann
Signature	

p.p. Julia Heinzmann

REV 09_00 /2021-02-08