

**Certificate of Analysis
GM501 Collagenase**

Batchnumber	COLL-2206
Production date	07 October 2022
Expiry date	31 July 2023
Total volume (mL)	1000

	Results	Specifications
pH	7,35	7,20-7,70
Osmolality	279	270-310 mOsm/kg
Bacterial screen	PASSED	PASSED
Fungal screen	PASSED	PASSED
Sperm survival Test 4 hours exposure	97	≥ 80 %

Release date	24 October 2022
Released by	Vanessa Bierlich
Signature	

P.P. 

REV 09_00 /2021-02-08