

**Certificate of Analysis
GM501 Collagenase**

Batchnumber	COLL-2306
Production date	09 August 2023
Expiry date	31 May 2024
Total volume (mL)	860

	Results	Specifications
pH	7,3	7,20-7,70
Osmolality	282	270-310 mOsm/kg
Sterility test (according to the current Ph.Eur.2.6.1 USP <71>)	No growth	No growth
HSSA (% motility compared with control after 4 h exposure to test medium)	92	≥ 80 %

Release date	05 September 2023
Released by	Nicole Geiger
Signature	

pp. N. Geiger

REV 10_00 /2023-07-03