

**Certificate of Analysis
GM501 Collagenase**

Batchnumber COLL-2502
 Production date 01 August 2025
 Expiry date 31 May 2026
 Total volume (ml) 3120

Product was sterilized by aseptic processing techniques

Parameter	Results	Specifications
pH	7,32	7,20-7,70
Osmolality	281	270-310 mOsm/kg
Sterility test (according to the current Ph.Eur.2.6.1 USP <71>)	No growth	No growth
HSSA (% motility compared with control after 4 h exposure to test medium)	100	≥ 80 %

Release date 18 August 2025
 Released by Joana Sturm
 Signature

