

**Certificate of Analysis
GM501 Collagenase**

Batchnumber COLL-2503
 Production date 02 December 2025
 Expiry date 30 September 2026
 Total volume (ml) 1560

Product was sterilized by aseptic processing techniques

Parameter	Results	Specifications
pH	7,36	7,20-7,70
Osmolality	283	270-310 mOsm/kg
Sterility test (according to the current Ph.Eur.2.6.1 USP <71>)	No growth	No growth
HSSA (% motility compared with control after 4 h exposure to test medium)	100	≥ 80 %

Release date 07 January 2026
 Released by Nicole Geiger
 Signature

pp. N. Geiger

REV 11_00 / 2023-09-15