

**Certificate of Analysis  
GM501 Collagenase**

Batchnumber COLL-2603  
 Production date 15 July 2026  
 Expiry date 30 April 2027  
 Total volume (ml) 1562

Product was sterilized by aseptic processing techniques

| Parameter                                                                 | Results   | Specifications  |
|---------------------------------------------------------------------------|-----------|-----------------|
| pH                                                                        | 7,36      | 7,20-7,70       |
| Osmolality                                                                | 281       | 270-310 mOsm/kg |
| Sterility test (according to the current Ph.Eur.2.6.1 USP <71>)           | No growth | No growth       |
| HSSA (% motility compared with control after 4 h exposure to test medium) | 93        | ≥ 80 %          |

Release date 30 July 2026  
 Released by Dr. René Richarz  
 Signature

*p.p. Richarz*

REV 11\_00 / 2023-09-15