

**Certificate of Analysis  
Micropipettes - ICSI**

Batchnumber 20200108  
 Production date 08 January 2020  
 Expiry date 31 January 2025

**Article No.**  
 IC-5-30°-s  
 IC-5-35°-be  
 IC-5-35°-s

|  | <b>Results</b> | <b>Specifications</b> |
|--|----------------|-----------------------|
| Packaging  | PASSED         | PASSED                |
| Sterilization  | PASSED         | PASSED                |
| Mouse Embryo Assay<br>(1-cell, blastocysts after 96 hours) | 100            | ≥ 80 %                |

Release date 11 February 2020  
 Released by Signature Dr. Julia Heinzmann

*p.p. Julia Heinzmann*