

**Certificate of Analysis
Micropipettes - Holding**

Batchnumber 20200114
 Production date 14 January 2020
 Expiry date 31 January 2025

Article No.

HP-110-30°
 HP-120-35°

	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	100	≥ 80 %

Release date 28 February 2020

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann

REV 03_00 / 2020-01-27