

**Certificate of Analysis  
Micropipettes - ICSI**

Batchnumber	20200513
Production date	13 May 2020
Expiry date	31 May 2025
<b>Article No.</b>	
IC-5-35°-s	

	<b>Results</b>	<b>Specifications</b>
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	95	≥ 80 %

Release date	03 July 2020
Released by Signature	Dr. Julia Heinzmann

*p.p. Julia Heinzmann*