

**Certificate of Analysis  
Micropipettes - Biopsy**

Batchnumber 20200729  
 Production date 29 July 2020  
 Expiry date 31 July 2025

**Article No.**

BP-10-0°-be

	<b>Results</b>	<b>Specifications</b>
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	95	≥ 80 %

Release date 15 September 2020  
 Released by Dr. Julia Heinzmann  
 Signature

*p.p. Julia Heinzmann*