

Certificate of Analysis Micropipettes - Holding

Batchnumber	20200729
Production date	29 July 2020
Expiry date	31 July 2025

Article No.

HP-120-30°
HP-120-35°

	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay <small>(1-cell, blastocysts after 96 hours)</small>	95	≥ 80 %

Release date	15 September 2020
--------------	-------------------

Released by	Dr. Julia Heinzmann
-------------	---------------------

Signature

p.p. Julia Heinzmann