

**Certificate of Analysis
Micropipettes - Holding**

Batchnumber 20200916
 Production date 16 September 2020
 Expiry date 30 September 2025

Article No.

HP-100-30°
 HP-100-35°
 HP-120-30°

	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	90	≥ 80 %

Release date 12 October 2020

Released by Dr. Julia Heinzmann

Signature

p.p. Julia Heinzmann

REV 03_00 / 2020-01-27