

**Certificate of Analysis
Micropipettes - ICSI**

Batchnumber 20200928

Production date 28 September 2020

Expiry date 30 September 2025

Article No.

IC-5-30°-s

IC-5-35°-s

	Results	Specifications
Packaging	PASSED	PASSED
Sterilization	PASSED	PASSED
Mouse Embryo Assay (1-cell, blastocysts after 96 hours)	90	≥ 80 %

Release date 19 October 2020

Released by Signature Dr. Julia Heinzmann

p.p. Julia Heinzmann